

MULTISPECTRUM ROTATIONAL STATES DISTRIBUTION THERMOMETRY

RICCARDO GOTTI, MARCO LAMPERTI, DAVIDE GATTI, *Department of Physics - Institute of Photonics and Nanotechnology, Politecnico di Milano, Lecco, Italy*; SZYMON WOJTEWICZ, *Institute of Physics, Faculty of Physics, Astronomy and Informatics, Nicolaus Copernicus University, Torun, Poland*; THOMAS PUPPE, YURIY MAYZLIN, *Frequency Combs Research and Development, TOPTICA Photonics AG, Gräfelfing, Germany*; BIDOOR ALSAIF, *Clean combustion research center, King abdullah university for science and technology, Thuwal, Saudi arabia*; JULIAN ROBINSON-TAIT, FELIX ROHDE, RAFAL WILK, PATRICK LEISCHING, WILHELM KAENDERS, *Frequency Combs Research and Development, TOPTICA Photonics AG, Gräfelfing, Germany*; PAOLO LAPORTA, MARCO MARANGONI, *Dipartimento di Fisica, Politecnico di Milano, Milano, Italy*.

We introduce Multispectrum Rotational states Distribution Thermometry (MRDT) as a new optical method for primary thermometry that relies on the global fitting of multiple molecular absorption lines of the same band at different pressures. This allows leveraging the temperature-dependence of the Doppler width and also of the distribution of line intensities across the ro-vibrational band, provided a sufficiently accurate line-strength model is available. We give a preliminary demonstration of the method with a comb-locked frequency-swept cavity-ring-down spectrometer operated on the $3\nu_1 + \nu_3$ band of CO₂ located around 1577 nm, which stands out among other spectroscopic samples for the availability of several line intensity models of both experimental and theoretical origin. The spectra signal-to-noise ratio represents the main limitation to a combined uncertainty to 530 ppm, but the comparative analysis between different line-strength models shows promise to reduce the error budget to 33 ppm. As compared to Doppler-broadening-thermometry, an advantage of the approach is the reduced impact of a wrong modelling of the absorption line-shapes. In a reversed approach, MRDT can be applied on a gas of known temperature to set an upper limit to the accuracy of a given line intensity model.