WIDE-BANDWIDTH COMB-ASSISTED SPECTROSCOPY IN THE FINGERPRINT REGION AND APPLICATION TO THE ν_1 FUNDAMENTAL BAND OF $^{14}\mathrm{N}_2$ $^{16}\mathrm{O}$

BIDOOR ALSAIF, Clean combustion research center, King abdullah university for science and technology, Thuwal, Saudi arabia; MARCO LAMPERTI, DAVIDE GATTI, PAOLO LAPORTA, Dipartimento di Fisica, Politecnico di Milano, Milano, Italy; MARTIN FERMANN, Laser Research, IMRA AMERICA, Inc, Ann Arbor, MI, USA; AAMIR FAROOQ, Clean combustion research center, King abdullah university for science and technology, Thuwal, Saudi arabia; OLEG LYULIN, Laboratory of Theoretical Spectroscopy, Institute of Atmospheric Optics, Tomsk, Russia; ALAIN CAMPARGUE, UMR5588 LIPhy, Université Grenoble Alpes/CNRS, Saint Martin d'Hères, France; MARCO MARANGONI, Dipartimento di Fisica, Politecnico di Milano, Milano, Italy.

Most spectroscopic data available in databases such as HITRAN are retrieved from FTIR measurements and suffer from uncertainties at the MHz level. Much more accurate data, by up to four orders of magnitude, can be achieved using an optical frequency comb to calibrate the frequency axis of a cw laser source. Actually, in the mid-infrared region, at least beyond 5 μ m, the only available commercial solution for a widely tunable cw laser is represented by extended cavity quantum cascade lasers (EC-QCLs), whose locking to an optical frequency comb has been so far inhibited by a large amount of frequency noise, leading to linewidths of about 20 MHz a .

In this work we overcome this limitation and describe a spectrometer that relies on the frequency locking of an EC-QCL tunable in the 7.55-8.2 μm range to a 1.9 μm Tm fiber comb b . It is applied to the first comb-calibrated direct characterisation of the ν_1 fundamental band of N_2O , specifically of nearly 70 lines in the $1240-1310~{\rm cm}^{-1}$ range, from P(40) to R(31). The spectroscopic constants of the upper state are derived from a fit of the line centers with an average rms uncertainty of $4.8 \times 10^{-6}~{\rm cm}^{-1} (144~{\rm kHz})$. The coupling of the spectrometer to a high-finesse optical cavity to the purpose of enhancing its sensitivity and addressing weaker absorbers, is also discussed.

^aKnabe K., Williams P. A., Giorgetta F., Armacost C. M., Crivello S., Radunsky M., and Newbury N., Opt. Express 20, 12432-12442 (2012)

^bLamperti M., Alsaif B., Gatti D., Fermann M., Farooq A., and Marangoni M., Sci. Rep. 8,1292 (2018)